

DECLARATION OF CONFORMITY

TO COUNCIL DIRECTIVE 93/42/EEC: 2007/47/EC CONCERNING MEDICAL DEVICES

Manufacturer: **Name: JiangSu YuYue Medical Equipment & Supply CO., LTD.**
Address: Yunyang Industrial Park, Danyang City, Jiangsu Province , China . 212300

European Representative:
Name: Shanghai International Holding Corp.GmbH(Europe)
Address: Eiffestrasse 80, 20537 Hamburg Germany

Product Name: Nebulizer
Model: M102

Classification (MDD, Annex IX): Ila (Rule 11)
Conformity Assessment Route: **MDD Annex V.3**

We herewith declare that the above mentioned products meet the transposition into national law, the provisions of Council Directive 93/42/EEC: 2007/47/EC concerning medical devices. All supporting documentations are retained under the premises of the manufacturer.

DIRECTIVES

General applicable directives:
Medical Device Directive: COUNCIL DIRECTIVE 93/42/EEC: 2007/47/EC concerning medical devices (MDD 93/42/EEC).

STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.

Notified Body: TÜV SÜD Product Service GmbH, Ridlerstrabe 65, 80339, München, Germany

Identification number: CE0123

(EC) Certificate(s): G2 055329 0025 Rev.00

Start of CE Marking: Date CE mark was affixed:2020-03-30

Expire date of the Certificate: 2024-05-26

Place, Date of Issue: DanYang, JiangSu , P.R.CHINA 2020-02-14

Signature: 

Name: Bill Wang

Position : Management Representative

JIANGSU YUYUE MEDICAL EQUIPMENT & SUPPLY CO. LTD
江苏鱼跃医疗设备股份有限公司


LIST OF EU HARMONISED AND INTERNATIONAL STANDARDS

S/N	Document No.	Edition	Title
1	93/42/EEC	2007/47/EC	Medical device directives of EU
2	ISO 13485	2003	Medical devices-Quality management systems-Requirements for regulatory purposes
3	ISO 14971	2000	Medical devices - Application of risk management to medical devices
4	EN ISO 10993-1	2003	Biological evaluation of medical devices – Part 1: Evaluation and testing
5	EN ISO 10993-5	1999	Biological evaluation of medical devices – Part 5: Tests done in vitro cytotoxicity
6	EN ISO 10993-10	2002	Biological evaluation of medical devices – Part 10: Tests for irritation and delayed-type
7	EN 980	2003	Graphical symbols for use in the labelling of medical devices
8	EN 1041	1998	Information supplied by the manufacturer with medical devices
9	EN 60601-1	1990 +A1:1993 +A2: 1995 +A3: 1996	Medical electrical equipment – Part 1: General requirements for safety (IEC 60601-1:1998)
10	EN 60601-1-2	2001	Medical electrical equipment – Part 1-2: General requirements for safety – Collateral standard: Electromagnetic compatibility – Requirements and tests
11	EN ISO 8359	1996	Oxygen concentrators for medical use – Safety requirements
12	EN 60601-1-4	1996 +A: 1999	Medical electrical equipment – Part 1-4: General requirements for safety – Collateral standard: Programmable electrical medical systems

JIANGSU TUBE MEDICAL EQUIPMENT & SUPPLY CO., LTD
江苏鱼跃医疗设备股份有限公司
张明